

Product Change Notification

Issue Date: Mar,30, 2018

Product Name: MEDP-HD-BR-R10

Status: Running change

This notice is to inform you of changes to the MEDP-HD-BR-R10.

Reason for Change

MEDP-HD-BR-R10 will be EOL and MEDP-HD-BR-R11 will comply with the latest EMC standard for IEC60601-1-2(Ed.4.0).

Description of Change

Add ESD IC

Affected Model

Affected Model	New Version
MEDP-HD-BR-R10	MEDP-HD-BR-R11

Affected by This Change

- | | | | |
|---|-------------------------------------|-------------------------------------|----------------------------------|
| <input type="checkbox"/> Specifications | <input type="checkbox"/> Dimensions | <input type="checkbox"/> Appearance | <input type="checkbox"/> Package |
| <input checked="" type="checkbox"/> Certification | <input type="checkbox"/> Label | <input type="checkbox"/> BIOS | <input type="checkbox"/> Driver |

For more details please refer to the latest datasheet or visit <http://www.ieiworld.com>.

Best Regards,

David

Project Manager

IEI Integration Corp.